

Biolog OmniLog Incubator Reader
Specifications

| Item | |
|------------------------------|---|
| Incubation Temperature Range | 22° to 45° C |
| Temperature Consistency | ±2° C in the tray chamber |
| Incubation Humidity Range | Ambient |
| Test Capacity | 50 microplates |
| Reading Cycle Time | Up to 25 trays with 50 microplates in 15 mins |